

Semi-automated blood coagulation analyser

CA-104



Registered under Act 737 Registration No: IVDA1335722-87555

Sysmex Asia Pacific Pte Ltd
Tel +65 6221-3629 Fax +65 6221-3687
www.sysmex-ap.com

Sysmex India Pvt. Ltd Tel +91 (22) 6112-6666 www.sysmex.co.in

PT Sysmex Indonesia Tel +62 (21) 3002-6688 Fax +62 (21) 3002-6699 **www.sysmex.co.id**

Sysmex (Malaysia) Sdn Bhd Tel +60 (3) 5637-1788 Fax +60 (3) 5637-1688 **www.sysmex.com.my**

Sysmex New Zealand Ltd Tel +64 (9) 630-3554 www.sysmex.co.nz

Sysmex (Thailand) Co., LtdTel +66 (2) 032-2536 Fax +66 (2) 116-5396 **www.sysmex.co.th**

Sysmex Philippines Inc.Tel +63 (2) 621-2460 Fax +63 (2) 621-2432 **www.sysmex.com.ph**

Sysmex Vietnam Co., LtdTel +84 (8) 3997-9400 Fax +84 (8) 3997-9405 **www.sysmex.com.vn**

Sysmex Vietnam Co., Ltd (Hanoi Branch) Tel +84 (4) 3776-7020 Fax +84 (4) 3776-7022 **www.sysmex.com.vn**

Sysmex Australia Pty Ltd Tel +61 (2) 9016-3040 www.sysmex.com.au

SAP/HAEMOSTASIS/CA-104/30042019/V3.0

www.sysmex-ap.com

Incubate sample

Timer countdown

Dispense sample



Aspirate start reagent





Dispense start reagent, measurement is started automatically

Result printed out automatically



Reagent system

Thromborel S, Dade Innovin

APTT • Dade Actin, Dade Actin FS, Dade Actin FSL, Pathromtin SL

TT • Thromboclotin, Test Thrombin Reagent

Multifibren U

CA-104 Your Straightforward, Compact Coagulation Analyser

Ideal for laboratories that require minimum manual intervention in performing routine coagulation assays. It is designed to simplify coagulation testing; making it more standardized and reproducible

- Coagulation testing is made easy with built-in timer, temperature control, and auto-start and timing of measuring process
- Easy to use with pre-programmed methods
- Objective detection of clot formation
- · Automatic light intensity adjustment according to the turbidity of the plasma, making it possible to measure icteric or lipemic plasma reliably
- Sophisticated standard curve tool to maintain accuracy and precision
- Increase throughput with short incubation time
- A technology that requires only half the sample and reagent volume for testing as compared to manual method



Turbodensitometric Measuring Principle

A light beam passes through the cuvette containing the test plasma onto a photo detector. Any change in the intensity of the transmitted light, i.e., light increase or decrease, is converted into an electric signal

The stirrer mix the reagent and plasma in the cuvette. And at the same time creates a small whirl through the mixer movement to ensure even the smallest fibrin clot is formed in front of the photo detector

Once the start reagent is added, the lamp intensity automatically adjusts up or down according to the turbidity of the sample

This stirring action and optical measurement constitute the basic features of the turbodensitometric measuring principle

